



**Branch:**

Regional Inspectorate Brno, Okružní 31, 638 00 Brno,  
tel. +420 545 555 111, fax. +420 545 555 183

## CERTIFICATE OF CONFORMITY

**No. 6015-CC-A299-19**

List 1 z 1 listu

**In accordance:** with Directive 2014/32/EU of the European Parliament and of the Council on the harmonization of the laws of the Member States relating to the making available on the market of measuring instruments (implemented in Czech Republic by Government Order No. 120/2016 Coll.).

**Measuring instrument:** Inductive flow meter

**Manufacturer:** Hydromess s.r.o.  
Holandská 878/2  
639 00 Brno  
Czech Republic

**Type:** F401 ACQUAMAG

**EC-type exam. certificate:** SK 11-MI001-SMU023      **Issued by:** SMÚ      **Valid until:** 1. 12. 2021

**Serial number:** 192423, 192428

**Specifications:** Nominal diameter: DN 80  
Software version: F401.34  
Checksum CRC16: 3392  
Continuous flow  $Q_3$ : 160 m<sup>3</sup>/h  
Ratio  $Q_3/Q_1$ : R250  
Temperature class: T50  
Class of water pressure: MAP 16  
Classes: M1; E1; B  
Positions: H,V  
Flow profile sensitivity classes: U5, D3  
Measurement: one direction


**Accuracy class:** 2

**Harmonized standard or normative document:** ISO 4064

**Date of issue:** 2019, September 16



**Certificate approved by:**

  
Ing. Radovan Wiecek